



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994

OPW FUEL MANAGEMENT SYSTEMS
 6900 Santa Fe Drive
 Hodgkins IL 60525
 Akram Elmofty 708 387 8241

CALIBRATION

Valid until: February 28, 2018

Certificate Number: 2794.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Mechanical

Parameter/Equipment	Range	CMC ² (±)	Comments
Magnetostrictive Probe Calibration with a Laser Device	Probe Length (914 to 3785) mm	0.074 mm	Stated uncertainty is for the calibration process. The maximum allowable linearity error for a calibrated probe does not exceed 0.90 mm

¹ This laboratory is not normally available for commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.



Accredited Laboratory

A2LA has accredited

OPW FUEL MANAGEMENT SYSTEMS

Hodgkins, IL

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSLI Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).



Presented this 24th day of February 2016.

A handwritten signature in blue ink, appearing to read "J. C. Bunt".

Senior Director of Quality & Communications
For the Accreditation Council
Certificate Number 2794.01
Valid to February 28, 2018

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.